

<p>Bath Iron Works A GENERAL DYNAMICS COMPANY</p>	<h1>Appendix G</h1>		
<p>Title: SUPPLEMENTAL REQUIREMENTS FOR GENERAL MATERIAL DDG 51 ARLEIGH BURKE CLASS and DDG 1000 ZUMWALT CLASS</p>		<p>Rev. S</p>	<p>Date 3/20/08</p>

APPENDIX G HOME	REVISION HISTORY
<p>SDRL CODE: QCOD, CALIBRATION CERTIFICATE:</p> <ol style="list-style-type: none"> a. The Seller shall accomplish the requirement of ANSI Z540-1-1994, Calibration Laboratories and Measuring and Test Equipment-General Requirements. b. Two legible copies of the Certificate of Calibration, Figure 4, shall be provided with each packing list. Certificate shall be completed and signed by a responsible representative of the Seller. c. Two legible copies of the Calibration Data Sheets shall also be provided by Seller with each packing list to indicate functions tested, nominal and measured values, and calibration tolerances. <hr/> <p style="text-align: center;"> <u>CALIBRATION CERTIFICATE</u> <u>Printable .pdf Version</u> </p> <p>I hereby certify that the instrument(s) listed below have been examined, tested and found to be within the manufacturer's published specification when compared with standard instruments.</p> <p>Calibration of all standard instruments used on performance of the above tests is traceable to the National Bureau of Standards. The standards and instruments used in the calibration are supported by a calibration system that meets or exceeds the requirements of ANSI Z540-1-1994.</p> <p style="text-align: right;">Signed _____ Title _____ Company _____</p> <p>Mfr: _____ Item Part # _____ Purchase Order # _____ Item Serial # _____</p>	

Line Item _____ Calibration Data _____
BIW Catalog # _____ Calibration Due Data _____
Item Nomen. _____

National Bureau of Standards Test No. _____

Figure (4) Calibration Certificate Figure (4)